

Certificate of Compliance

**Manufacturer:**

Phoenix Medical Systems (P) Ltd
DP-42, SIDCO Industrial Estate,
Thirumudivakkam, Chennai, 600 044, India

Authorized Representative:

Obelis s.a
Bd. General Wahis 53, 1030 Brussels,
Belgium

This is to certify that the technical documentation for the product:

Product Name : Mobility Aid

Model No/Type : SmartCane

Technical File No : TF-SC-A

Product description:

Smart cane device is an electronic travel aid for persons with visual impairment when used in conjunction with a white cane for mobility and navigation.

The device is used along with a White Cane. It produces ultrasound waves for obstacle detection within 3m of the user. The vertical detection range extends from knee-height (about 0.27 times the total height) to the head height of the person.

Complies with the requirements applicable to it

The manufacturer's technical documentation, as required for Class I devices, has been assessed as meeting the Essential Requirements and relevant provisions of the Directive 93/42/EEC, as amended by 2007/47/EC, on Medical Devices for CE Marking.

Certificate issued by:

A handwritten signature in black ink that reads 'C. Y. Won'.

ChanYo, Won
Regional Technical Director
UL Health Sciences

Certificate Number: 12CA29641
Certificate expiry: 2017-04-15



Any significant changes in design or construction of the product, or amendments to the Directive or Standards referenced above may render this receipt invalid. The product liability rests with the manufacturer or his representative in accordance with Council Directive 93/42/EEC.

While all due care and skill was exercised in carrying out this assessment, UL Health Sciences accept responsibility only for proven gross negligence. This certificate is issued for the date that the technical documentation was reviewed. Conformance to all regulatory requirements is the sole responsibility of the manufacturer including the manufacture and quality control of the products. This is not legal document and cannot be used as such. This certificate remains the property of UL Health Sciences to whom it must be returned on request.